

English

Intended use

The Roche **cobas b 101** is an in vitro diagnostic test system designed to quantitatively determine the C-reactive protein (CRP) in human capillary whole blood and serum, EDTA K2/K3 and lithium heparin anticoagulated whole blood and plasma by photometric measurement. Measurement of CRP is of use for the evaluation of inflammatory disorders and associated diseases, infection and tissue injury. The system is intended for use in point-of-care (PoC) settings such as pharmacies, physician offices, physician office laboratories, clinics and hospitals, and clinical laboratory settings.

Summary

Most tissue-damaging processes such as infections, inflammatory diseases and malignant neoplasms are associated with a major acute phase response of the C-reactive protein (CRP) and other acute phase reactants (e.g. AAT, AAGP, C3C, C4, HAPT). The CRP response frequently precedes clinical symptoms, including fever. In healthy individuals CRP is a trace protein with a range up to 5 mg/L. After onset of an acute phase response the serum CRP concentration rises rapidly and extensively. Alterations are detectable within 6 to 8 hours and the peak value is reached within 24 to 48 hours. Levels of up to a thousandfold the normal value are associated with severe stimuli such as myocardial infarction, major trauma, surgery, or malignant neoplasms. CRP activates the classical complement pathway. CRP has a half-life of only a few hours, making it an ideal tool for clinical monitoring. Postoperative monitoring of CRP levels of patients indicates either the normal recovery process (decreasing levels to normal) or unexpected complications (persisting high levels). Measuring changes in the concentration of CRP provide useful diagnostic information about how acute and how serious a disease is. It also allows the assessment of complications during the disease and judgements about the disease genesis. Persistence of a high CRP concentration is usually a grave prognostic sign which generally indicates the presence of an uncontrolled infection. CRP determination may replace the classical determination of Erythrocytes Sedimentation Rate (ESR), due to its prompt response to changes in disease activity and its good correlation to ESR.^{1,2,3,4}

Test principle

The erythrocytes of the capillary or venous blood sample are separated from the plasma by centrifugation. Then, the plasma sample is diluted with HEPES buffer and transferred into a reaction chamber where it is mixed with CRP antibody-latex reagent. The CRP in the diluted plasma binds with the CRP antibody on the latex particle. The concentration of CRP is calculated as a function of the changed absorbance measured at 525 nm and 625 nm which is in relation to the amount of agglutination.^{5,6,7}

Reagents

One test contains:

HEPES buffer: 1.79 mg

Anti-human CRP antibody (goat) Latex-conjugate: 41.84 µg

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

Reagent handling

Carefully tear open the foil pouch at the tear notch until one side is open.

Discard the disc if the foil pouch is found open or damaged, or if the disc is damaged, or the desiccant is missing, or loose desiccant particles or any other dirt or particles especially at the blood application zone are found.

Use **cobas CRP Control** in the same way as a blood sample.

Storage and stability

Store at 2-30 °C until the expiration date printed on the pouch. Do not freeze. If stored in a refrigerator, allow the disc to warm up in the closed pouch for at least 20 minutes before use. Once the pouch is opened, use the disc within 20 minutes. Protect the disc from direct sunlight. Do not store opened pouches in a refrigerator.

Specimen collection and preparation

For specimen collection and preparation only use suitable tubes or collection containers.

Use fresh human capillary whole blood or serum, EDTA K2/K3 or lithium heparin anticoagulated whole blood or plasma.

Do not use other anticoagulants or other additives.

For EDTA K2/K3 and lithium heparin anticoagulated whole blood and plasma samples, test within 8 hours of sample collection if stored at room temperature. If stored in the refrigerator plasma samples may be used up to 14 days and whole blood samples up to 3 days. Frozen serum and plasma samples stored at -20 °C may be used up to 30 days. Freeze only one time. Mix sample thoroughly before use. Do NOT use frozen whole blood to avoid the risk of hemolysis.

The marking on the disc shows where to apply the sample. If samples are used from a venipuncture or control material, use a standard pipette or dropper to form a drop. The disc is self-filling. Do not push the sample into the disc. Do not use syringes. Assure that the disc is free from blood outside the sample application zone and the hinge cover.

Sample volume: 12 µL

Sample stability on disc

① After sample application, the disc must be inserted immediately (in ≤ 120 seconds). Please follow the instructions in the **cobas b 101** Operator's Manual.

Assay

Instructions for use

- Wash hands with soap. Warm water helps to stimulate the blood flow. Rinse the fingers extensively. Dry hands.
- Disinfect the fingertip by wiping three times the area to be lanced with a cotton swab or sterile gauze pad impregnated with 70 %-100 % isopropanol emollient free or 70 %-100 % ethanol emollient free; repeat the procedure with a second cotton swab or sterile gauze pad impregnated with 70 %-100 % isopropanol emollient free or 70 %-100 % ethanol emollient free, then dry with a cotton swab or sterile gauze pad.
- Prick the patient's finger by applying a single-use disposable lancing device (e.g. Accu-Chek Safe-T-Pro Plus). Make sure to follow the according lancing device instructions for obtaining a blood sample.
- Wipe off the first drop of blood with a swab.
- With the imprinted side of the disc facing upwards, position the disc's suction point above the drop of blood. The disc is self-filling.
- Apply blood and observe that it has filled the marked area. Check the sample volume: turn the disc on its backside. The area marked in blue has to be filled completely with blood. Do not push the blood into the disc.
- Press hinge cover down firmly to close the disc.
- Assure that the disc is free from blood outside the sample application zone and the hinge cover.
- Insert the disc into the **cobas b 101** instrument. Close the lid.
- The measurement starts automatically.

For more details, please refer to the **cobas b 101** Quick Reference Guide or **cobas b 101** Operator's Manual.

Materials provided

- [REF] 08024669190, **cobas CRP Test**, 10 tests

Materials required (but not provided)

- Single use disposable lancing device (e.g. Accu-Chek Safe-T-Pro Plus)
- [REF] 08024723190, **cobas CRP Control**
- [REF] 06378668190, **cobas b 101** instrument
- Optical check disc
- General laboratory equipment (e.g., sample transfer pipette for venous blood or alcohol wipes for disinfection of the finger)

cobas CRP Test



Calibration

This method has been standardized against the ERM DA 472/IFCC reference material. Each disc lot of the **cobas CRP Test** is traceable to ERM DA 472/IFCC reference material.^{8,9}

The instrument automatically reads in the lot-specific calibration data from the barcode information printed on the disc, eliminating the need for calibration by the user.

Quality control

For quality control, use **cobas CRP Control**.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

QC info disc

Every **cobas CRP Control** kit contains a lot-specific QC info disc for quality control. This QC info disc contains the target values and ranges for the **cobas CRP Test**.

The instrument display prompts the user to insert the QC info disc. The **cobas b 101** instrument reads the disc providing the lot specific target ranges.

Display of results

At the end of the automatic determination, the **cobas b 101** instrument shows the result within 3-4 minutes. The concentration of CRP will be displayed in mg/L or in mg/dL.

Limitations - interference

Hematocrit levels between 20 % and 60 % do not affect results.

Criterion: Recovery within ± 10 % of initial values at CRP concentrations of 10.0 mg/L and 40.0 mg/L.

Hemolysis: No significant interference up to a hemoglobin concentration of 500 mg/dL.

Icterus:¹⁰ No significant interference up to a conjugated/unconjugated bilirubin concentration of 50 mg/dL.

Lipemia (Intralipid):¹⁰ No significant interference up to an Intralipid and Triglyceride concentration of 750 mg/dl.

Glycemia:¹⁰ No significant interference up to a glucose level of 1000 mg/dL. A fasting sample is not required.

Rheumatoid factors: No significant interference up to 1200 IU/mL.

Drugs: No interference was found at therapeutic concentrations using common drug panels.¹¹

Drug interferences are measured based on recommendations given in CLSI guidelines EP07 and EP37 and other published literature. Effects of concentrations exceeding these recommendations have not been characterized.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

Limits and ranges

Measuring range

3.0-400 mg/L or 0.30-40.0 mg/dL

Expected values

Adults: < 5.0 mg/L (< 0.5 mg/dL)^{2,12}

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

Specific performance data

Representative performance data on the instruments are given below. Results obtained in individual laboratories may differ.

Precision

Precision was determined using controls in a CLSI EP5-A3 protocol.

Precision was measured with 3 lots of **cobas CRP Test** using 5 different serum samples at the medical decision points and 2 **cobas CRP Control** solution levels over 21 days with 2 runs per day and duplicate

measurements per run and specimen. The following results were obtained for a representative lot:

Sample	Repeatability			Intermediate precision		
	Mean mg/L	SD mg/L	CV %	Mean mg/L	SD mg/L	CV %
Sample Healthy	5.1	0.13	2.5	5.1	0.17	3.3
Sample Cut off	10.0	0.23	2.3	10.0	0.24	2.4
Sample Decision	39.9	0.93	2.3	39.9	0.98	2.5
Sample Acute	93.4	1.62	1.7	93.4	1.84	2.0
Sample Acute High	351	7.99	2.3	351	8.42	2.4
Control Level 1	9.7	0.29	2.9	9.7	0.30	3.1
Control Level 2	39.2	0.79	2.0	39.2	1.09	2.8

Method comparison

A comparison experiment using 3 different reagent lots with serum samples measured with **cobas CRP Test** on the **cobas b 101** instrument (y) and CRP-latex X2 "Seiken" NX reagent on the **cobas c 501** analyzer (x) gave the following correlation for a representative lot (Weighted Deming regression method).^{13,14}

Sample size: 140

Slope: 1.00

Intercept: 0.0934

Pearson r: 0.998

Mean bias in the range 3.0-200 mg/L: 0.5 %

Mean bias in the range > 200-400 mg/L: 3.5 %

Bias at 5.0 mg/L: 1.9 %

Bias at 10.0 mg/L: 0.9 %

References

- Guidance for Industry and FDA Staff Review Criteria for Assessment of C-Reactive Protein (CRP), High Sensitivity C-Reactive Protein (hsCRP) and Cardiac C-Reactive Protein (cCRP) Assays; 2005, p. 1246.
- Aguilar FJ, Ferreira-Júnior M, Sales MM, et al. C-reactive protein: clinical applications and proposals for a rational use. Rev Assoc Med Bras Jan-Feb 2013, Vol. 59, pp. 85-92.
- van Leeuwen MA and van Rijswijk MH. Acute phase proteins in the monitoring of inflammatory disorders. Baillieres Clin Rheumatol., 1994, Vol. 8, pp. 531-52.
- Gabay C. and Kushner I. Acute-Phase Proteins and Other Systemic Responses to Inflammation. N Engl J Med, 1999, Vol. 340, pp. 448-454.
- Senju O, Takagi Y, Uzawa R, et al. A new immuno quantitative method by latex agglutination-application for the determination of serum C-reactive protein (CRP) and its clinical significance. J Clin Lab Immunol., 1986, Vol. 19, pp. 99-103.
- Price CP, Trull AK, Berry D, et al. Development and validation of a particle-enhanced turbidimetric immunoassay for C-reactive protein. J of Immunol Methods, 1987, Vol. 99, pp. 205-211.
- Eda S, Kaufmann J, Roos W, et al. Development of a New Microparticle-Enhanced Turbidimetric. J. Clin Lab Anal., 1998, Vol. 12, pp. 137-144.
- Baudner S, Bienvenu J, Blirup-Jensen S, et al. The certification of a matrix reference material for immunochemical measurement of 14 human serum proteins CRM470. Report EUR 15243 EN. Commission of the European Communities, 1993, pp. 1-186.
- Richtlinie der Bundesärztekammer zur Qualitätssicherung. Deutsches Ärzteblatt. Sep 19, 2014, pp. 1583-1618.
- Kroll MH and Elin RJ. Interferences with clinical laboratory analyses. Clin Chem, 1994, Vol. 40, pp. 1996-2005.
- Sonntag O. and Scholer A. Drug interference in clinical chemistry: recommendation of drugs and their concentrations to be used in drug interference studies. Ann Clin Biochem, 2001, Vol. 38, pp. 376-385.
- Pepys MB and Hirschfield GM. C-reactive protein: a critical update. J. Clin. Invest, 2003, Vol. 111, pp. 1805-1812.

cobas CRP Test




- 13 Martin RF. General Deming Regression for Estimating Systematic Bias and Its Confidence Interval in Method-Comparison Studies. Clinical Chemistry, 2000, Vol. 46, pp. 100-104.
- 14 Johnson R. Assessment of Bias with Emphasis on Method Comparison. Clin Biochem Rev 2008; Vol. 29, pp. S37-S42.

For further information, please refer to the appropriate Operator's Manual for the instrument concerned, and the Method Sheets of all necessary components.

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see dialog.roche.com for definition of symbols used):

CONTENT	Contents of kit
SYSTEM	Analyzers/Instruments on which reagents can be used
REAGENT	Reagent
CALIBRATOR	Calibrator
	Volume after reconstitution or mixing
GTIN	Global Trade Item Number

COBAS, COBAS B, and COBAS C are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2019, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com

